

IIIFHD - Report Period: APR 2016
Status Of Reported Tests

STATUS	COUNT	PERCENT
Aborted	5	9.259
Valid	49	90.741
Total Reported Tests	54	100.000

IIIFHD - Report Period: APR 2016
Itemized Lost Test Report

Obs	Date	TESTKEY	REASON
1	20151012	ACC-17064	Test Terminated At Start Up Due To Connecting Rod Failure.
2	20151030	ACC-17164	Test Terminated At Test Hour 20 Due To 10 Hour Purge Not Returned To Engine.
3	20151123	ACC-17355	Test Terminated At Test Hour 55 Due To Oil Filter Block Temperature Control Problems.
4	20160205	ACC-17802	Test Terminated At Test Hour 54 Due To Spun Bearing.
5	20160324	ACC-17948	Test Terminated At Test Hour 17 Due To The Inability To Maintain Load.

IIIFHD - Report Period: APR 2016
Replicate Database Report

Components of Replicate Database	APR 2016	All Periods
Number of Tests	.	18
Formulations	.	9
Labs	.	4
Stands	.	9

IIIFHD - Report Period: APR 2016
by-Formulation Pooled s and Reproducibility of Replicate Data

Report Period	Parameter	Degrees of Freedom	Pooled s	R
All	PVIS	9	0.33071	0.92598
All	PVIS (Without SA)	9	0.06376	0.17853

IIIFHD - Report Period: APR 2016
Pass/Fail

Report Period	API Category	N	Number Passed	% Passed
APR 2016	All Categories, All Parameters	49	45	91.8367
APR 2016	CJ-4 - All Parameters	49	45	91.8367
APR 2016	CJ-4, PVIS	49	45	91.8367
APR 2016	CJ-4, PVIS (Without SA)	49	45	91.8367
All	All Categories, All Parameters	906	800	88.3002
All	CJ-4 - All Parameters	906	800	88.3002
All	CJ-4, PVIS	906	800	88.3002
All	CJ-4, PVIS (Without SA)	906	800	88.3002




Monitoring Agency

6555 Penn Avenue, Pittsburgh, PA 15206

<https://acc-ma.org>

IIIFHD
Status of Reported Tests






Monitoring Agency

6555 Penn Avenue, Pittsburgh, PA 15206

<https://acc-ma.org>

IIIFHD

Candidate Precision as Represented by Pooled S





Monitoring Agency

6555 Penn Avenue, Pittsburgh, PA 15206

<https://acc-ma.org>

IIIFHD

Candidate Precision as Represented by Pooled S

