

Status Of Reported Tests

STATUS	COUNT	PERCENT
Aborted	1	14.286
Cancelled	1	14.286
Invalid	2	28.571
Valid	3	42.857
Total Reported Tests	7	100.000

Itemized Lost Test Report

Obs	Date	TESTKEY	REASON
1	20231006	ACC-34344	Test Failed To Meet Soot Window
2	20231024	ACC-34322	Test Failed To Meet Soot Requirements
3	20231201	ACC-33635	Test Aborted Due To Early Viscosity Increase.
4	20240108	ACC-35019	Test Terminated At Sponsor Request

Replicate Database Report

Components of Replicate Database	APR 2024	All Periods
Number of Tests	.	18
Formulations	.	9
Labs	.	3
Stands	.	8

by-Formulation Pooled s and Reproducibility of Replicate Data

Report Period	Parameter	Degrees of Freedom	Pooled s	R
All	VISI38	9	1.30741	3.66075
All	RV48	9	0.47125	1.31950
All	VISI38 (Without SA)	9	1.25138	3.50387
All	RV48 (Without SA)	9	0.46976	1.31534
All	RV2	9	0.24502	0.68605
All	RV2FNL	9	0.24502	0.68605

Pass/Fail

Report Period	API Category	N	Number Passed	% Passed
APR 2024	All Categories, All Parameters	3	2	66.667
APR 2024	CI-4 - All Parameters	3	2	66.667
APR 2024	CI-4, RV2	3	2	66.667
APR 2024	CI-4, RV2FNL	3	2	66.667
APR 2024	CI-4, RV48	3	2	66.667
APR 2024	CI-4, RV48 (Without SA)	3	2	66.667
APR 2024	CI-4, VISI38	3	2	66.667
APR 2024	CI-4, VISI38 (Without SA)	3	3	100.000
All	All Categories, All Parameters	624	395	63.301
All	CI-4 - All Parameters	624	395	63.301
All	CI-4, RV2	624	467	74.840
All	CI-4, RV2FNL	624	459	73.558
All	CI-4, RV48	624	419	67.147
All	CI-4, RV48 (Without SA)	624	432	69.231
All	CI-4, VISI38	624	581	93.109
All	CI-4, VISI38 (Without SA)	624	583	93.429

ACC Monitoring Agency

203 Armstrong Drive, Freeport, PA 16229, USA

<https://www.acc-ma.org>
412-365-1000


Services provided by:



A Program of ASTM International

T8E

Status of Reported Tests



ACC Monitoring Agency

203 Armstrong Drive, Freeport, PA 16229, USA

<https://www.acc-ma.org>
412-365-1000

Services provided by:



A Program of ASTM International

T8E

Candidate Precision as Represented by Pooled S

